



Protecting your  
laboratory environment


**LABCONCO**<sup>®</sup>

LABCONCO CORPORATION 8811 Prospect Ave, Kansas City, MO 64132  
(816) 333-8811, Fax (816) 363-0130, (800) 821-5525

## INSTALLATION OF VENTS ON PROTECTOR SOLVENT STORAGE CABINETS

1. Remove cover plates on both the inside and outside back of cabinet.
2. Install vent connection parts as shown in view below.

NOTE: Barrel plugs are furnished if it is desired to close up vents.



## **Warranty-**

We are committed to providing our customers with quality equipment and service after the sale. Part of this objective involves keeping you informed of changes and new product additions.

Labconco provides a warranty on all parts and factory workmanship. The warranty includes areas of defective material and workmanship, provided such defect results from normal and proper use of the equipment.

The warranty for all Labconco products will expire one year from date of installation or two years from date of shipment from Labconco, whichever is sooner, except the following:

- Purifier® Delta™ Series Biological Safety Cabinets, which carry a three-year warranty from date of installation or four years from date of shipment from Labconco, whichever is sooner.
- Carts carry a lifetime warranty.
- Glassware is not warranted from breakage when dropped or mishandled.

This limited warranty covers parts and labor, but not transportation and insurance charges. In the event of a warranty claim, contact Labconco Corporation or the dealer who sold you the product. If the cause is determined to be a manufacturing fault, the dealer or Labconco Corporation will repair or replace all defective parts to restore the unit to operation. Under no circumstances shall Labconco Corporation be liable for indirect, consequential, or special damages of any kind. This statement may be altered by a specific published amendment. No individual has authorization to alter the provisions of this warranty policy or its amendments. Lamps and filters are not covered by this warranty. Damage due to corrosion or accidental breakage is also not covered.

**WARNING:** various federal, state or local regulations may govern the disposal and/or emission of substances used in connection with this equipment. All users of this equipment are urged to become familiar with any regulations that apply in the user's area concerning the dumping of waste materials in or upon water, land or air and to comply with such regulations.

## **Shipping Claims-**

If a shipment is received in visibly damaged condition, be certain to make a notation on the delivering carrier's receipt and have his agent confirm the damage on your receipt. Otherwise, the damage claim may be refused.

If concealed damage or pilferage is discovered, notify the carrier immediately and retain the entire shipment intact for inspection. Interstate Commerce Commission rules requires that the claim be filed with the carrier within 15 days after delivery.

**NOTE: Do not return goods.** Goods returned without prior authorization will not be accepted. Labconco Corporation and its dealers are not responsible for shipping damage. The recipient must file claims directly with the freight carrier. If authorization has been received to return this product, by accepting this approval, the user assumes all responsibility and liability for biological and chemical decontamination and cleansing. Labconco reserves the right to refuse delivery of any products that do not appear to have been properly cleaned and/or decontaminated prior to return.

**Contacting Labconco-**

If you have any questions that are not addressed in these instructions, or if you need technical assistance, please contact Labconco's Sales Information Department at either 1-(800) 821-5525, and Service Information at (800)-522-7658 or (816) 333-8811, between the hours of 7:00 a.m. and 6:00 p.m. Central Standard Time.

Visit the Internet at <http://www.labconco.com>

Or

e-mail: [labconco@labconco.com](mailto:labconco@labconco.com)